



भारतीय प्रौद्योगिकी संस्थान भिलाई  
जी.ई.सी. कैपस, सेजबहार, रायपुर - ४९२०१५  
छत्तीसगढ़, भारत

Indian Institute of Technology Bhilai  
G.E.C. Campus, Sejbahar, Raipur - 492015  
Chhattisgarh, India

Date: 13-09-2022

### CORRIGENDUM-1

This is in reference to the tender no. IITBh/Goods/CIF/2022-23/065 Dated: 07/09/2022 for Supply and Installation of Spectral Domain Based High-Speed Optical Coherence Tomography (OCT) System at IIT Bhilai.

The following changes have been made in the tender documents of Supply and Installation of Spectral Domain Based High-Speed Optical Coherence Tomography (OCT) System at IIT Bhilai:

| S No.                            | Item                                                                                           | Existing Technical Specification                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   | Updated/ Modified Technical Specification                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    |
|----------------------------------|------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1                                | Specifications of computer system<br><br>Page No. 13<br>Section - V                            | The specifications of computer system: It should be with Windows 10 Professional 64-bit operating system having Intel Processor (CPU: Intel 8 Core 3 GHz or better, DDR RAM 32 GB or better, suitable HDD or better, Video graphics card, suitable DVD writer, Gigabit Ethernet, USB ports 2.0 and 3.0, Fire wire. Keyboard, mouse, mouse pad, 30 inch or better LED monitor. A suitable online UPS (not less than 2 KVA) to run the complete system (full OCT system + Computer CPU + Monitor etc) at least for 15 min in case of electricity power-cut. One portable 5TB or better capacity external hard drive should also be provided for the storage of data. | The specifications of computer system: It should be with Windows 10 Professional 64-bit operating system having Intel Processor (CPU: Intel 8 Core 3 GHz or better, DDR RAM 32 GB or better, suitable HDD or better, Video graphics card, 128 GB or better pendrive/memory stick USB3.0, Gigabit Ethernet, USB ports 2.0 and 3.0, Fire wire. Keyboard, mouse, mouse pad, 30 inch or better LED monitor. A suitable online UPS (not less than 2 KVA) to run the complete system (full OCT system + Computer CPU + Monitor etc) at least for 15 min in case of electricity power-cut. One portable 5TB or better capacity external hard drive should also be provided for the storage of data. |
| <b>Additional Payment terms:</b> |                                                                                                |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |
| 2                                | Payment:<br><br>Point No. 7<br><br>Page No. 11<br><br>Inclusion of Foreign Payment LC Facility | -                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  | 7.3 For foreign currency payments: 90% of the payment will be made through irrevocable Letter of Credit (LC). Balance 10% on successful installation, commissioning and acceptance by IITBh through LC, any loss due to fluctuation in foreign exchange rates will be at beneficiary account. The LC operating charges inside India will be to IITBh account and those outside India will be to beneficiary account. The LC confirmation charges & LC amendment charges, if requested/ required, will be to beneficiary account.                                                                                                                                                             |

All other terms and conditions of the tender document will remain the same

Deputy Registrar  
(Stores and Purchase)